

New England Biolabs Certificate of Analysis

Product Name: *SacII*
Catalog Number: *R0157S*
Concentration: *20,000 U/ml*
Unit Definition: *One unit is defined as the amount of enzyme required to digest 1 µg of pXba DNA in 1 hour at 37°C in a total reaction volume of 50 µl.*
Lot Number: *10014128*
Expiration Date: *07/2020*
Storage Temperature: *-20°C*
Storage Conditions: *50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA*
Specification Version: *PS-R0157S/L v1.0*

SacII Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0157SVIAL	SacII	10014129	Pass
B7204SVIAL	CutSmart® Buffer	10010634	Pass
B7024SVIAL	Gel Loading Dye, Purple (6X)	10011266	Pass


Assay Name/Specification	Lot # 10014128
Endonuclease Activity (Nicking) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of supercoiled pBR322 DNA and a minimum of 60 Units of SacII incubated for 4 hours at 37°C results in <20% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³ H] E. coli DNA and a minimum of 100 units of SacII incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Ligation and Recutting (Terminal Integrity) After a 20-fold over-digestion of pXba DNA with SacII, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with SacII.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of pXba DNA and a minimum of 60 Units of SacII incubated for 16 hours at 37°C results in a DNA pattern free of	Pass

Assay Name/Specification	Lot # 10014128
detectable nuclease degradation as determined by agarose gel electrophoresis.	

This product has been tested and shown to be in compliance with all specifications.



Tony Spear-Alfonso
Production Scientist
06 Jun 2018



Michael Tonello
Packaging Quality Control Inspector
10 Jul 2018