


## New England Biolabs Certificate of Analysis

**Product Name:** Random Primer 6  
**Catalog Number:** S1230S  
**Packaging Lot Number:** 10076481  
**Expiration Date:** 06/2023  
**Storage Temperature:** -20°C  
**Storage Conditions:** Supplied as a lyophilized triethylammonium salt  
**Specification Version:** PS-S1230S v1.0


Random Primer 6 Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
S1230SVIAL	Random Primer 6	10076482	Pass

Assay Name/Specification	Lot # 10076481
<b>Electrophoretic Pattern (Oligonucleotide)</b> The banding pattern of Random Primer 6 on a 15% TBE-Urea gel is evaluated against a control lot for expected size, sharpness, and relative intensity.	<b>Pass</b>

This product has been tested and shown to be in compliance with all specifications.



Michael Kuska  
Production Scientist  
22 Jun 2020



Josh Hersey  
Packaging Quality Control Inspector  
22 Jun 2020