

New England Biolabs Certificate of Analysis

Product Name: NEBNext® Ultra™ RNA Library Prep Kit for Illumina®
 Catalog Number: E7530L
 Lot Number: 10025002
 Expiration Date: 02/2020
 Storage Temperature: -20°C
 Specification Version: PS-E7530S/L v1.0

NEBNext® Ultra™ RNA Library Prep Kit for Illumina® Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E7431AAVIAL	Nuclease-free Water	10019131	Pass
E7427AAVIAL	NEBNext® Second Strand Synthesis Reaction Buffer, 10X	10019132	Pass
E7425AAVIAL	NEBNext® Second Strand Synthesis Enzyme Mix	10019127	Pass
E7424AAVIAL	RNase Inhibitor, Murine	10019125	Pass
E7423AAVIAL	ProtoScript® II Reverse Transcriptase	10019133	Pass
E7422AAVIAL	Random Primers	10019124	Pass
E7421AAVIAL	NEBNext® First Strand Synthesis Reaction Buffer	10019123	Pass
E7373AAVIAL	Blunt/TA Ligase Master Mix	10019130	Pass
E7372AAVIAL	NEBNext® End Repair Reaction Buffer	10019129	Pass
E7371AAVIAL	NEBNext® End Prep Enzyme Mix	10019128	Pass
E6625AAVIAL	NEBNext® Q5® Hot Start HiFi PCR Master Mix	10019134	Pass

Assay Name/Specification	Lot # 10025002
Functional Testing (Library Construction, RNA) Each set of reagents is functionally validated and compared to a previous lot through construction of libraries made from commercially available RNA, using the kit's minimum and maximum input requirements. Libraries made from previous and current lots for both input RNA amounts are sequenced together on the same Illumina® flow cell and compared across various metrics including library yield, individual transcript abundance correlations (low vs. high input, old lot vs. new lot), 5'-3' transcript coverage, and fraction of reads mapping to a reference.	Pass
* Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEBNext® Ultra™ RNA Library Prep Kit for Illumina® and meet the designated specifications.	Pass

This product has been tested and shown to be in compliance with all specifications.



Christine Sumner

Christine Sumner
Production Scientist
03 Oct 2018

Michael Tonello

Michael Tonello
Packaging Quality Control Inspector
03 Oct 2018