

## New England Biolabs Certificate of Analysis

**Product Name:** *Endonuclease VIII*  
**Catalog Number:** *M0299L*  
**Concentration:** *10,000 U/ml*  
**Unit Definition:** *One unit is defined as the amount of enzyme required to cleave 1 pmol of a 34-mer oligonucleotide duplex containing a single AP site in a total reaction volume of 10 µl in 1 hour at 37°C in 1X Endonuclease VIII Reaction Buffer containing 10 pmol of fluorescently labeled oligonucleotide duplex.*  
**Lot Number:** *10013336*  
**Expiration Date:** *06/2019*  
**Storage Temperature:** *-20°C*  
**Storage Conditions:** *10 mM Tris-HCl, 250 mM NaCl, 0.1 mM EDTA, 50 % Glycerol, (pH 8.0 @ 25°C)*  
**Specification Version:** *PS-M0299S/L v1.0*

Endonuclease VIII Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
M0299LVIAL	Endonuclease VIII	10013337	Pass
B0299SVIAL	Endonuclease VIII Reaction Buffer	0011606	Pass

Assay Name/Specification	Lot # 10013336
<b>Protein Purity Assay (SDS-PAGE)</b> Endonuclease VIII is ≥ 95% pure as determined by SDS-PAGE analysis using Coomassie Blue detection.	Pass
<b>Non-Specific DNase Activity (16 Hour)</b> A 50 µl reaction in Endonuclease VIII Reaction Buffer containing 1 µg of Lambda-HindIII DNA and a minimum of 30 units of Endonuclease VIII incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
<b>Exonuclease Activity (Radioactivity Release)</b> A 50 µl reaction in Endonuclease VIII Reaction Buffer containing 1 µg of a mixture of single and double-stranded [ <sup>3</sup> H] E. coli DNA and a minimum of 10 units of Endonuclease VIII incubated for 4 hours at 37°C releases <0.5% of the total radioactivity.	Pass

This product has been tested and shown to be in compliance with all specifications.

*Lauren Higgins*

---

Lauren Sears Higgins  
Production Scientist  
20 Jun 2018

*Michael Tonello*

---

Michael Tonello  
Packaging Quality Control Inspector  
20 Jun 2018