

## New England Biolabs Certificate of Analysis


**Product Name:** *Phusion® High-Fidelity PCR Master Mix with GC Buffer*  
**Catalog Number:** *M0532L*  
**Concentration:** *2 X Concentrate*  
**Packaging Lot Number:** *10194240*  
**Expiration Date:** *08/2024*  
**Storage Temperature:** *-20°C*  
**Specification Version:** *PS-M0532S/L v2.0*  
**Composition (1X):** *1 X Phusion® GC Buffer, 0.2 mM dATP, 0.2 mM dCTP, 0.2 mM dGTP, 0.2 mM dTTP, 20 units/ml Phusion® High-Fidelity DNA Polymerase*

Phusion® High-Fidelity PCR Master Mix with GC Buffer Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
M0532SVIAL	Phusion® High-Fidelity PCR Master Mix with GC Buffer	10180959	Pass
B0515AVIAL	DMSO	10194246	Pass

Assay Name/Specification	Lot # 10194240
<b>PCR Amplification (20 kb Lambda DNA, Master Mix)</b> A 50 µl reaction in 1X Phusion® High-Fidelity PCR Master Mix with GC Buffer and 1.0 µM primers containing 10 ng Lambda DNA for 22 cycles of PCR amplification results in the expected 20 kb product.	Pass
<b>PCR Amplification (7.5 kb Human Genomic DNA, Master Mix)</b> A 50 µl reaction in 1X Phusion® High-Fidelity PCR Master Mix with GC Buffer and 1.0 µM primers containing 50 ng Human Genomic DNA for 30 cycles of PCR amplification results in the expected 7.5 kb product.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.

  
Lea Antonopoulos  
Production Scientist  
13 Mar 2023

  
Michael Tonello  
Packaging Quality Control Inspector  
18 Aug 2023