

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: 3'-Desthiobiotin-GTP

Catalog Number: N0761S
Concentration: 5 mM
Lot Number: 10052316
Expiration Date: 08/2022
Storage Temperature: -20°C

Specification Version: PS-N0761S v1.0

Composition (1X): Supplied in Ultrapure water as a Triethylammonium salt (pH 7.0).

3'-Desthiobiotin-GTP Component List			
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result
N0761SVIAL	3'-Desthiobiotin-GTP	10052317	Pass

Assay Name/Specification	Lot # 10052316
Concentration (A260) The concentration of 3'-Desthiobiotin-GTP is between 5 and 5.5 mM as determined by	Pass
UV absorption at 260 nm.	
Endonuclease Activity (Nicking)	Pass
A 50 µl reaction in NEBuffer 4 containing 1 µg of supercoiled PhiX174 DNA and a minimum of 5 nmoles of 3'-Desthiobiotin-GTP incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	
Functional Testing (Capping, 3'DTB-GTP)	Pass
A 20 µl reaction in Capping Buffer containing 0.5 mM 3'-Desthiobiotin-GTP, 0.1 mM SAM, 62.5 ng of a 25 mer RNA transcript, 20 units of Vaccinia Capping Enzyme and 0.05 units of Yeast Inorganic Pyrophosphatase incubated for 30 minutes at 37°C results in >90% capped transcript as determined by gel electrophoresis.	
Physical Purity (HPLC)	Pass
3'-Desthiobiotin-GTP is ≥ 95% pure as determined by HPLC analysis.	
RNase Activity (Extended Digestion)	Pass
A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA	
and a minimum of 5 nmoles of 3'-Desthiobiotin-GTP is incubated at 37°C. After incubation for 16 hours, >90% of the substrate RNA remains intact as determined by	
gel electrophoresis using fluorescent detection.	



N0761S / Lot: 10052316

Page 1 of 2



This product has been tested and shown to be in compliance with all specifications.

Bo Wu

Production Scientist 13 Aug 2019 Michael Tonello

Packaging Quality Control Inspector

14 Aug 2019

N0761S / Lot: 10052316

Page 2 of 2