

New England Biolabs Certificate of Analysis

Product Name: pUC19 Vector
Catalog #: N3041S/L
Concentration: 1,000 µg/ml
Unit Definition: N/A
Lot #: 0381407
Assay Date: 07/2014
Expiration Date: 07/2016
Storage Temp: -20°C
Storage Conditions: 10 mM Tris-HCl (pH 8.0), 1 mM EDTA
Specification Version: PS-N3041S/L v1.0
Effective Date: 08 Jul 2014

Assay Name/Specification (minimum release criteria)	Lot #0381407
A260/A280 Assay - The ratio of UV absorption of pUC19 Vector at 260 and 280 nm is between 1.8 and 2.0.	Pass
DNA Concentration (A260) - The concentration of pUC19 Vector is between 1000 and 1050 µg/ml as determined by UV absorption at 260 nm.	Pass
Electrophoretic Pattern (Plasmid) - The banding pattern of pUC19 Vector on a 1.2% agarose gel is evaluated against a control lot for sharpness and relative intensity as determined by gel electrophoresis using Ethidium Bromide.	Pass
Non-Specific DNase Activity (DNA, 16 hour) - A 50 µl reaction in 1X NEBuffer 2 containing 5 µg of pUC19 Vector incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
Restriction Digest (Linearization) - A 50 µl reaction in CutSmart™ Buffer containing 5 µg of pUC19 Vector DNA and 20 units of XbaI incubated for 1 hour at 37°C produces > 95% linearization resulting in a single band of approximately 2686 bp as determined by agarose gel electrophoresis.	Pass



Authorized by
Derek Robinson
08 Jul 2014



Inspected by
Vanessa Mathieu-Sheltry
02 Jul 2014

