


New England Biolabs Certificate of Analysis

Product Name: *Fast DNA Ladder*
Catalog #: *N3238S*
Concentration: *25 µg/ml*
Unit Definition: *N/A*
Lot #: *0021401*
Assay Date: *01/2014*
Expiration Date: *01/2016*
Storage Temp: *4 °C*
Storage Conditions: *0.002 % Xylene cyanol , 5 mM Tris-HCl (pH 8.0), 5 mM EDTA , 5 % Glycerol*
Specification Version: *PS-N3238S v1.0*
Effective Date: *25 Jun 2014*

Assay Name/Specification (minimum release criteria)	Lot #0021401
DNA Concentration (A260) - The concentration of Fast DNA Ladder is between 25 and 27.5 µg/ml as determined by UV absorption at 260 nm.	Pass
Electrophoretic Pattern (Marker) - The banding pattern of Fast DNA Ladder on a 1.2% agarose gel shows discrete, clearly identifiable bands at each band of the marker, when stained with Ethidium Bromide at a concentration of 0.5 µg/ml.	Pass
Non-Specific DNase Activity (DNA, 16 hour) - A 50 µl reaction in 1X NEBuffer 2 containing 1.25 µg of Fast DNA Ladder incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass



Authorized by
Derek Robinson
25 Jun 2014



Inspected by
Vanessa Mathieu-Sheltry
24 Jul 2014

