

New England Biolabs Certificate of Analysis

Product Name: BspEI
Catalog Number: R0540S
Concentration: 10,000 U/ml
Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA (dam) in 1 hour at 37°C in a total reaction volume of 50 µl.
Lot Number: 10027093
Expiration Date: 10/2020
Storage Temperature: -20°C
Storage Conditions: 300 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 500 µg/ml BSA
Specification Version: PS-R0540S/L v1.0

BspEI Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0540SVIAL	BspEI	10027092	Pass
B7203SVIAL	NEBuffer™ 3.1	10021113	Pass


Assay Name/Specification	Lot # 10027093
<p>Blue-White Screening (Terminal Integrity) A sample of LITMUS38i vector linearized with a 10-fold excess of BspEI, religated and transformed into an E. coli strain expressing the LacZ beta fragment gene results in <1% white colonies.</p>	Pass
<p>Endonuclease Activity (Nicking) A 50 µl reaction in NEBuffer 3.1 containing 1 µg of supercoiled PhiX174 DNA and a minimum of 10 Units of BspEI incubated for 4 hours at 37°C results in <20% conversion to the nicked form as determined by agarose gel electrophoresis.</p>	Pass
<p>Exonuclease Activity (Radioactivity Release) A 50 µl reaction in NEBuffer 3.1 containing 1 µg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 100 units of BspEI incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.</p>	Pass
<p>Ligation and Recutting (Terminal Integrity) After a 10-fold over-digestion of Lambda dam- DNA with BspEI, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated</p>	Pass

Assay Name/Specification	Lot # 10027093
fragments, >95% can be recut with BspEI. Non-Specific DNase Activity (16 Hour) A 50 µl reaction in NEBuffer 3.1 containing 1 µg of Lambda dam- DNA and a minimum of 50 units of BspEI incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass

This product has been tested and shown to be in compliance with all specifications.



Jianying Luo
Production Scientist
01 Nov 2018



Michael Tonello
Packaging Quality Control Inspector
20 Dec 2018