

New England Biolabs Certificate of Analysis

Product Name: TspMI
Catalog Number: R0709S
Concentration: 5,000 U/ml
Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg of pUCAdeno plasmid DNA in 1 hour at 75°C in a total reaction volume of 50 µl.
Lot Number: 10020625
Expiration Date: 02/2019
Storage Temperature: -20°C
Storage Conditions: 300 mM NaCl, 20 mM Tris-HCl (pH 8.0), 1 mM DTT, 1 mM EDTA, 50% Glycerol, 0.10% Triton X-100, 200 µg/ml BSA
Specification Version: PS-R0709S/L v1.0

TspMI Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0709SVIAL	TspMI	10020135	Pass
B7204SVIAL	CutSmart® Buffer	10013536	Pass


Assay Name/Specification	Lot # 10020625
<p>Endonuclease Activity (Nicking) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of supercoiled PhiX174 DNA and a minimum of 5 Units of TspMI incubated for 4 hours at 75°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.</p>	Pass
<p>Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 50 units of TspMI incubated for 4 hours at 75°C releases <0.1% of the total radioactivity.</p>	Pass
<p>Ligation and Recutting (Terminal Integrity) After a 10-fold over-digestion of pUCAdeno DNA with TspMI, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 25°C. Of these ligated fragments, >75% can be recut with TspMI.</p>	Pass
<p>Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of pUCAdeno DNA and a minimum of 5 Units of TspMI incubated for 16 hours at 75°C results in a DNA pattern free of</p>	Pass

Assay Name/Specification	Lot # 10020625
detectable nuclease degradation as determined by agarose gel electrophoresis.	

This product has been tested and shown to be in compliance with all specifications.



Tony Spear-Alfonso
Production Scientist
21 Jun 2018



Michael Tonello
Packaging Quality Control Inspector
31 Aug 2018