

New England Biolabs Certificate of Analysis

Product Name: *Oligo d(T)25 Magnetic Beads*
Catalog Number: *S1419S*
Concentration: *5 mg/ml*
Packaging Lot Number: *10124280*
Expiration Date: *07/2024*
Storage Temperature: *4°C*
Storage Conditions: *0.02 % NaN₃, 0.05 % Tween®20, 1 X PBS, (pH 7.4 @ 25°C)*
Specification Version: *PS-S1419S v2.0*

Oligo d(T)25 Magnetic Beads Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
S1419SVIAL	Oligo d(T) ₂₅ Magnetic Beads	10113849	Pass

Assay Name/Specification	Lot # 10124280
Non-Specific DNase Activity (16 hour, Buffer) A 50 µl reaction in Oligo d(T)25 Magnetic Bead Storage Buffer containing 1 µg of PhiX174-HaeIII DNA incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
Functional Testing (mRNA Isolation) Oligo d(T)25 Magnetic Beads were equilibrated and incubated with freshly prepared eukaryotic cell lysate for direct mRNA isolation. The beads were washed and the mRNA eluted. The eluate was evaluated on an Agilent Bioanalyzer and the enriched poly(A)+ RNA contains ≤15% rRNA.	Pass
Binding Capacity (Magnetic Beads) Oligo d(T)25 Magnetic Beads (500 µg) were equilibrated and incubated with 100 µl of 67 µg/ml rA30 for 15 minutes at 25°C, then washed and the rA30 eluted. Binding capacity was determined to be >5 µg of rA30 per mg of beads.	Pass
RNase Activity (Extended Digestion) A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA and a minimum of 1 µl of Oligo d(T)25 Magnetic Beads is incubated at 37°C. After incubation for 16 hours, >90% of the substrate RNA remains intact as determined by gel electrophoresis using fluorescent detection.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.


Michael Spfoviero
Production Scientist
04 Nov 2021


Michael Tonello
Packaging Quality Control Inspector
04 Nov 2021