


New England Biolabs Certificate of Analysis

Product Name: Monarch® PCR and DNA Cleanup Kit
 Catalog Number: T1030S
 Lot Number: 0091807
 Expiration Date: 01/2020
 Storage Temperature: 25°C
 Specification Version: PS-T1030S/L v1.0

Assay Name/Specification	Lot # 0091807
Functional Testing (DNA Recovery and Purity) Twenty-four Monarch® DNA Cleanup Columns are tested with 1 µg of 1 Kb Ladder resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.	Pass
Functional Testing (Fragment Purification and Restriction Digestion) A linear pLA3 fragment is added to 8 Monarch® DNA Cleanup Columns, eluted and digested with DraIII-HF® resulting in ≥90% conversion to two-fragment pattern observed by gel electrophoresis in ≥75% of the samples.	Pass

This product has been tested and shown to be in compliance with all specifications.



Ana Egana
Production Scientist
27 Jun 2018



Heidi Church
Packaging Quality Control Inspector
27 Jun 2018